

 <p>IIT PALAKKAD</p>	<p>भारतीय प्रौद्योगिकी संस्थान पालक्काड Indian Institute of Technology Palakkad अहलिआ एकीकृत कैम्पस, कोज़िपारा Ahalia Integrated Campus, Kozhipara पालक्काड- 678557 Palakkad – 678 557</p>	<p>दूरभाषसंख्या/ Phone no: 04923 – 226 300/590/586</p> <p>ईमेल/ Email : purchase@iitpkd.ac.in</p>
---	--	---

Prof. Job Kurian
Registrar i/c

Ref : Differential Thermal Analyzer
Date : 28.11.2017

Open Tender No: IITPKD/CIF/DJ/074/2017

Due Date: 19.12.2017 @ 2.00 PM

Dear Sir/Madam,

On behalf of the **Indian Institute of Technology, Temporary campus, Palakkad, Quotations are invited for “Thermogravimetric – Differential Thermal Analyzer”**. The Specifications are given in the Annexure.

Pre-bid meeting – The Pre-bid meeting is scheduled to be held on 30.11.2017 at 10.00 AM in the Class Room, Academic Block, IIT Palakkad

Technical bid Opening: The Technical bid will be opened on 19.12.2017 at 2.00 PM at Conference Room, Academic Block, IIT Palakkad.

Instructions to the Bidder

- (i) **Preparation of Bids:** - The tenders should be submitted **under two-bid system (i.e.) Technical bid and Financial bid in separate envelopes**. The technical bid should consist of all technical details along with commercial terms and conditions. No prices should be included in technical bid. Financial Bid should indicate item – wise prices for the items mentioned in the technical bid. The technical and the financial bids should be put in separate covers and sealed. Both sealed covers should be put into a bigger cover. **Bids must either be spiral bound / stapled together. No loose sheets will be accepted. All pages must be numbered.**
- (ii) The Quotations duly sealed and superscribed on the envelope **with the reference No. and due date, should be addressed to the undersigned so as to reach on or before the due date stipulated above.**
- (iii) **Delivery of the tender:** - The tender shall be sent to the below-mentioned address either by post or by courier so as to reach this office before the due date and time specified in the Schedule. The offer/bid can also be dropped in the tender box on or before the due date and time specified in the schedule. **The tender box is kept in the office of the Academic Block, IIT Palakkad, Ahalia Integrated Campus, Kozhipara, Palakkad-678 557.**

- (iv) **Opening of the tender:** - The offer/Bids will be opened by a committee duly constituted for this purpose. The technical bids will be opened first and will be examined by a technical committee which will decide the suitability of the bid as per our specifications and requirements. The bidders will be invited for opening of Technical bids. **The Bidder's representative should carry authorization letter from their company empowering them to participate in the Pre-bid and tender opening meetings.** In respect of opening of financial bid, those bidders who are technically qualified only will be called.
- (v) **Prices:** - The price should be quoted in nett per unit (after breakup) and must include all packing and delivery charges indicated separately for each item. **The price indicated should be CIF/CIP Kochi.** The offer/bid should be exclusive of taxes and duties, which will be paid by the purchaser as applicable. The price should be quoted without custom duty. **The custom duty will be paid at concessional rate against duty exemption certificate.**
- (vi) **Agency Commission:** - Agency commission, if any, will be paid to the Indian agents in Rupees on receipt of the equipment and after satisfactory installation. Agency Commission will not be paid in foreign currency under any circumstances. The details should be explicitly shown in Tender even in the case of 'Nil' commission. The tenderer should indicate the percentage of agency commission to be paid to the Indian agent. **Terms of Delivery:** - The item should be supplied to our Institute as per Purchase order. The installation and commissioning should be completed as specified **by us in the attached schedule.**
- (vii) **Acceptance & Rejection:** IIT Palakkad reserves the full right to accept / reject any tender at **any** stage without assigning any reason.

Yours sincerely,

Registrar, IIT Palakkad

SCHEDULE

Important Conditions:

- 1) The due date for the submission of the tender is **19.12.2017 @ 2.00 PM**
- 2) The offers / bids should be submitted in two-bids systems (i.e.) Technical bid and financial bid. The Technical bid should consist of all technical details / specifications only. The Financial bid should indicate item-wise price for each item and it should contain all Commercial Terms and Conditions including Taxes (separately), transportation, packing & forwarding charges, installation, guarantee, payment terms, pricing terms etc. The Technical bid and financial bid should be put in separate covers **superscribed clearly as “Technical Bid” and “Financial bid”** and sealed. Both the sealed covers should be put in a bigger cover. Open Tender for “ **Thermogravimetric – Differential Thermal Analyzer**” should be written on the left side of the Outer bigger cover and sealed.
- 3) **EMD: -EMD should be at 2% (two percent) of the tender value quoted by the bidder.** The EMD should be enclosed with the financial bid which will not be opened for Technical evaluation. Enclosing the EMD in the Technical bid will automatically DISQUALIFY the tenderer. EMD should be in the form of DD in favour of “**Indian Institute of Technology Palakkad**” and payable at Palakkad”. The tender without EMD would be considered as UNSOLICITED and will be REJECTED. Photo/FAX copies of the Demand Draft/Banker’s pay orders will not be accepted. No interest will be paid for the EMD and the EMD will be refunded to the successful bidder on receipt of Performance Security.
- 4) **Performance Security:-** The successful bidder will be asked to submit Performance Security for an amount of 5% of the value of the contract/supply. The Performance Security may be furnished in the form of an Account Payee DD or FD Receipt from the commercial bank or Bank Guarantee from any nationalized bank of India. **Only after submission of Performance Security, Purchase Order/Work Order will be released / L.C will be opened.**
- 5) **Performance Security in the form of Bank Guarantee:-** In case the successful bidder is a foreign company and wishes to submit Performance Security in the form of Bank Guarantee, the Bank Guarantee should be routed through the Beneficiary Bank to the end user bank. Otherwise, the Indian Agent of the foreign vendor has to submit a Bank Guarantee from a Nationalized Bank of India.
- 6) The Bank Guarantee should remain valid for a period of sixty days beyond the date of completion of all contractual obligations of the supplier including the warranty obligations.

If an Indian agent is involved, the following documents must be enclosed:

- Foreign principal's proforma invoice indicating the commission payable to the Indian Agent and nature of after-sales service to be rendered by the Indian Agent.
 - Copy of the agency agreement with the foreign principal and the precise relationship between them and their mutual interest in the business.
- 7) The offer/bids should be sent only for a system or equipment that is available in the market and supplied to a number of customers. A list of customers in India and abroad with details must accompany the quotations. Quotations for a prototype machine will not be accepted.
 - 8) Original catalogue (not any photocopy) of the quoted model duly signed by the principals must accompany the quotation in the Technical bid. No prices should ever be included in the Technical bid.
 - 9) Compliance or Confirmation report with reference to the specifications and other terms & conditions should also be obtained from the principal.
- 10) Validity:** Validity of Quotation not less than 90 days from the due date of tender.
- 11) Delivery Schedule:-** The tenderer should indicate clearly the time required for delivery of the item. In case there is any deviation in the delivery schedule, liquidated damages clause will be enforced or penalty for the delayed supply period will be levied.
- 12) Risk Purchase Clause:-** In the event of failure of supply of the item/equipment within the stipulated delivery schedule, the purchaser has all the right to purchase the item/equipment from other sources on the total risk of the supplier under risk purchase clause.
- 13) Payment:-** No Advance payment will be made for Indigenous purchase. 100% Payment after supply and successful installation and commissioning and certification by the end user. In case of import supplies the payment will be made only through **100% Letter of Credit i.e. (50% payment will be released against shipping documents and 50% after successful installation and meeting acceptance criteria wherever the installation is being done).** Advance payment may be considered on submission of Bank Guarantee equal to the amount of advance payment.
- 14) On-site Installation:** - The equipment or machinery has to be installed and commissioned by the successful bidder within 15 to 20 days from the date of receipt of the **item at site of IIT Palakkad.**

- 15) Warranty/Guarantee:** - The offer should clearly specify the warranty or guarantee period for the machinery/equipment. Any extended warranty offered for the same has to be mentioned separately. (For more details please refer our Technical Specifications).
- 16) Late offer:** - The offers received after the due date and time will not be considered. The Institute shall not be responsible for the late receipt of Tender on account of Postal, Courier or any other delay.
- 17) Loading and unloading charges will be borne by the bidder/Supplier.**
- 18) Acceptance and Rejection:** - I.I.T. Palakkad has the right to accept the whole or any part of the Tender or portion of the quantity offered or reject it in full without assigning any reason.
- 19) Do not quote the optional items or additional items unless otherwise mentioned in the Tender documents / Specifications.**
- 20) Disputes and Jurisdiction:** - Any legal disputes arising out of any breach of contract pertaining to this tender shall be settled in the court of competent jurisdiction located within the city of Palakkad in Kerala.
- 21) All Amendments, time extension, clarifications etc., will be uploaded on the institute website only and will not be published in newspapers. Bidders should regularly visit the above website to keep themselves updated. No extension in the bid due date/ time shall be considered on account of delay in receipt of any document by mail.**

Acknowledgement:- It is hereby acknowledged that the tenderer has gone through all the conditions mentioned above and agrees to abide by them.

**SIGNATURE OF TENDERER
ALONG WITH SEAL OF THE
COMPANY WITH DATE**

Sl. No.	Specifications for Thermogravimetric – Differential thermal analyzer (TG-DTA)	
1	Furnace Temperature range	15°C to to 1600°C
2	Furnace type	Platinum furnace with integrated protective tube Thermocouple.
3	Heating Rate	0.1°C/min to 100°C/min from ambient to 1000 °C 0.1°C/min to 25 °C/min from 1000 to 1500 °C
4	Thermocouple	Pt-Pt/Rh Type R
5	Temperature Accuracy	Ambient to 1000°C ± 0.5°C or better 1000 – 1600 °C ± 1 °C or better
6	Calorimetric accuracy	Ambient to 1600 °C ± 5% or better
7	Atmosphere	<ul style="list-style-type: none"> • Static or Dynamic, Gaseous Atmosphere should include N₂, Ar, O₂, CO₂, air over full temperature range, with automatic gas flow control. • Assurance of performing measurement in H₂ flow (100% or diluted in inert gas) must be highlighted with a performance report.
8	Balance Design	Vertical Top loading
9	Balance precision	± 0.02 % or better
10	Sensor	Separate platform for sample and reference
11	Mass Sensitivity	0.2 µg or better
12	TGA load range	10 – 1000 mg
13	Sample Pan	180 µL, Platinum and Alumina pans with lids.
14	DTA range and sensitivity	±1 - ±1000 µV
15	Weight	Should not exceed 20 kg
16	DTA RMS Noise	0.007 °C
17	Scan rate	0.01 °C to 100 °C /min
18	Power supply	220 V 50 Hz
19	Flow rate of gases	Up to 0- 500 mL/min. Provisions to perform automated gas switching operation. Should be able to calibrate for all the gases mentioned.
20	Cooling rate	Must cool from 1600 °C to 100 °C in less than 45 min. The cooler must be supplied along with the equipment.
21	Software features	Automatic window based operating system for TGA, DTA, Activation energy calculation, curve smoothing, stress analysis and Kinetics. Capable of performing simultaneous measurement and analysis.
22	Gas mixer	<ul style="list-style-type: none"> • Compact stand-alone controller for delivering a custom gas environment for analytical instruments. Should include three gas inputs, each with its own mass flow controller (MFC) and a single outlet for delivery of a user defined custom gas mixture. • Should be able to mix Air, Argon, Carbon dioxide, Helium, Nitrogen, Oxygen • Gas input: 3 mass flow-controlled inputs (software controlled input ratio) • Gas output: 1 • Input pressure: Should with stand 15 – 90 psi • Flow rate: upto 600 mL/min • Interface: Should be interfaced with PC via USB, Gas flow controls must be ON/OFF • Weight should not exceed 1 kg • Dimensions should not exceed 100 mm X100 mm X 100mm

23	Certified Calibration Standards	Standard reference materials for the calibration of balance and temperature.
24	Mass Spectrometer	<p>An optional Mass spectrometer detector for the analysis of effluent gas must be quoted separately. The specifications of Mass spectrometer detector are</p> <ul style="list-style-type: none"> • Turbo pump with built-in Ion Gauge meter. • Interface: Direct type for capillary column • Temperature setting: 350 °C • Mass analyzer system: Quadrupole with pre filter • Mass range: 1-1200 amu • Scan rate: fully variable up to 12,500 amu/sec with 0.1 amu step size • Resolution: 1 amu • Additional spare filament 1 No. must be quoted separately apart from the standard supply of 2 Nos. • Filament voltage: user selectable from 10-100 eV e • Source temperature: 350 °C with independent heating • Sensitivity for EI with 1 pg of OctaFluoroNaphthalene (OFN) 800:1 • minimum for the entire mass range • Analyzer Quadrupole with pre-filter • Mass Stability +.0.1 m/z over 48 hours • EI Voltage 10 - 100 eV
25	Gas transfer line	<p>The transfer of samples from TG-DTA system to MS using a gas transfer line with following specifications</p> <ul style="list-style-type: none"> • Operating temperature of lines and gas sampling valves – 50-350 °C to prevent condensation of effluent gas. • Mass flow controller range 10 – 200 mL/min • Pump with membrane and flap valves with a capacity of 10nl/min and vacuum of 100mbar • Synthetic coarse dumper filter and fine filter of borosilicate glass • 230 V 50 Hz • Control option must be provided to choose between MS or in built TCD detectors or both.
26	Installation	Should be installed and tested free of cost (at our premises)
27	Details on number of units	<ul style="list-style-type: none"> • Completely functional TG-DTA – 1 unit • MS detector – 1 unit • Gas mixer – 1 unit • Platinum pan with lid – 5 Nos. • Alumina pan with lid – 20 Nos. • Gas flow meter – 1 unit • Chiller – 1 unit • Pump for MS – 1 unit • Pump for gas transfer – 1 unit • Additional thermocouple - 2 sets (1 set = 1 sample sensor + 1 reference sensor) should be quoted.
28	Computer	<ul style="list-style-type: none"> • Processor: Intel ® Core™ i5 -3550 CPU @ 3.30 GHz, 3.30 GHz Factory Installed Operating System: Microsoft® Windows® 7 Ultimate 32 bit Memory: 4 GB Hard Drive: 1 TB 7200rpm SATA Networking: Integrated Intel Gigabit Ethernet External I/O Ports: 4 USB 3.0 (4 rear), 4 USB 2.0 (2 front, 2rear), 1 Ethernet (RJ45), 1 serial (9-pin), 1 VGA out Expansion Slots: 2 PCI, 1 PCIe Removable Media: DVD Recordable Video: Integrated Graphics Mouse: Edge Mouse

		<p>- USB 2-button optical mouse with scroll Keyboard: USB Keyboard Audio: Integrated Audio.</p> <ul style="list-style-type: none"> • Should have suitable interface options to connect to TG-DTA equipment and Gas mixer to the PC.
29	Power supply	The equipment should be powered with 220 – 240 V, Single phase, 50 Hz supply. Indian type socket. Any other requirement must be communicated in the quote. Any additional power supply points must be mentioned.
30	Acceptance Criteria	<ul style="list-style-type: none"> • Successful installation of the equipment and inspection of all the accessories as per the P.O. • Demonstration of TGA, DTA and MS using standard samples • Demonstration of software features • Demonstration of calibration procedures • On-site training by factory trained engineer has to be arranged by the vendor. • Training should include operation of the equipment, calibration procedure, software applications, analysis, handling and maintenance of system for at least 20 man hours. • Complete system should have 3 years warranty including MS and Chiller. AMC of 3 years in addition to warranty period must be provided.
30	Delivery	10 – 12 weeks from the date of purchase order

Who can participate in the bid?

Only those bidders fulfilling the following criteria should respond to the tender.

1. The bidder should be either an Original Equipment Manufacturer (OEM) of **Thermogravimetric – Differential Thermal Analyzer** or should be an authorized representative (provide documentary proof) of an OEM.
2. The bidder should be a company registered under the Companies Act, 1956/2013 OR a Limited Liability Partnership /a registered partnership firm OR a sole-proprietorship entity. Appropriate Registration incorporation certificate must be submitted.
3. The bidder must have a registered office in Karnataka/Tamil Nadu/Telangana/Andhra Pradesh/Maharashtra or Kerala. Certificate of registration for the offices to be provided.
4. The bidder must also have a service center in Karnataka/Tamil Nadu/ Telangana/Andhra Pradesh or Kerala. Certificate of registration for the centers to be provided. Details about scope of service activities provided by the service centres must be provided. The contact details of the service engineers must be provided.
5. The bidder must be in existence in the business of **Thermogravimetric – Differential Thermal Analyzer** or allied fields for a minimum period of 5 previous financial years (before or since 01 April 2012). Documentary evidences of experience must be provided.
6. The bidder should have implemented orders of **Thermogravimetric – Differential Thermal Analyzer** worth exceeding INR 20 Lakhs during previous three financial years (01 April 2014 – 31 March 2017). Copies of the most recent purchase orders and certificates of successful implementation must be included. Copies of financial statements or evidence of turnover must be furnished.
7. The bidder should have documentary evidence of having supplied at least 5 Nos. of **Thermogravimetric – Differential Thermal Analyzer** to a Centrally Funded Technical Institution (e.g., IIT, NIT, IISc, IISER, etc.) in the recent past. The bidder must provide a certificate of satisfactory performance of the supplied equipment from the institute to which they have recently supplied. Contact details of the faculty-in-charge of the installed setup must also be provided.
8. The bidder must provide detailed specification of each equipment/item. Model numbers, data sheets and brochures must be included for each quoted equipment/accessories/item. Specifications corresponding to quoted model number must be available publicly via OEM's website for scrutiny. If not, bid can be disqualified on technical grounds.
- 9. Compliance sheet for the technical specification and OEM Brochure have to be attached along with the Technical bid. Vendor has to fill the compliance sheet and mention page number or reference number in OEM brochure. Unfilled / partially filled sheets lead to disqualification.**
- 10. Service and warranty for a minimum period of three years for the equipment must be provided. AMC for additional three years must be quoted separately.**