

CLARIFICATIONS AFTER PRE-BID MEETING □

Clarifications for IITPKD/CIF/RP/067/2017, Date: 29.11.2017

1. Technical Specification, Point 7: [C-V measurement functions]
Clear instructions to extract any of the mandatory parameters that do not have direct measurement support, from the measured parameters, must be provided.
2. Technical Specification, Point 14: [Text Fixture Unit]
The unit must have at least four ports to support the four SMUs.
3. Technical Specification, Point 16: [Control from Remote PC]
 GPIB interface is also acceptable if FLEX/VXI is not supported.
4. Technical Specification Point 19: [Cables]
In the interest of better signal integrity, the vendor may limit the length of half the number of required cables of each type to 1.5 m.
5. Who can participate in the bid?: Point 10
In case calibration is not covered under warranty and AMC, then the cost for calibration must be quoted separately.