

 IIT PALAKKAD	भारतीय प्रौद्योगिकी संस्थान पालक्काड Indian Institute of Technology Palakkad अहलिआ एकीकृत कैम्पस, कोज़िहपारा Ahalia Integrated Campus, Kozhipara पालक्काड- 678557 Palakkad – 678 557	दूरभाषसंख्या/ Phone no: 04923 – 226300/590/586 ईमेल/ Email : purchase@iitpkd.ac.in
--	---	---

Ref No: IITPKD/CIF/MP/070/2017

Dt: 30.11.2017

CORRIGENDUM

Sub: Corrigendum- reg.,

Ref No: 1. Tender no: IITPKD/CIF/MP/070/2017, Dated: 23.11.2017,

Due Date: 14.12.2017 AT 3.00 PM

2. Supply of Liquid Chromatography Mass Spectroscopy

DUE DATE OF THE TENDER: 21.11.2017 AT 3.00 PM

SI No	Existing Entry	To be read as
1.	Technical Specification Row No: 4 Scan speed : Should have the scans peed of 15,000 amu persec or above	Row No: 4 Scan speed : Should have the scanspeed of 17,000 amu persec or above
2.	Row No: 6 Polarity switching time: +ve/-ve polarity switching time between alternate scans should be less than 20msec	Row No: 6 Polarity switching time: +ve/-ve polarity switching time between alternate scans should be less than 25 msec
3.	Row No. 17 Pump Online membrane Vacuum Degasser – 3 channels	Row No. 17 Pump Online membrane Vacuum Degasser – 2 channels
4.	Row No. 19 UV Vis Detector A Semi micro flow cell (2.5 µL Volume, 5mm Cell path length, 12MPa pressure max) with temperature control should be supplied as standard.	Row No. 19 UV Vis Detector A Semi micro flow cell with temperature control should be supplied as standard.
5.	Row No. 20 Column Oven & Column C18 Column 150mm x 2.1mm, 1.9 micron column with guard column should be offered	Row No. 20 Column Oven & Column C18 Column 150mm x 2.1mm, 5 micron column with guard column should be offered C4 Column 250mm x 4.6 mm, 5 micron column with guard column should be offered
6.	Due date of the tender : 14.12.2017 @ 3 PM	Due date of the tender : 21.12.2017 @ 3.00 PM

Rest of the terms and conditions of IIT Palakkad will remain unchanged.

REGISTRAR, IIT PALAKKAD