

 IIT PALAKKAD	भारतीय प्रौद्योगिकी संस्थान पालक्काड Indian Institute of Technology Palakkad अहलिआ एकीकृत कैम्पस, कोज़िपारा Ahalia Integrated Campus, Kozhipara पालक्काड- 678557 Palakkad – 678 557	Phone no: 04923 – 226 300 /586 Email : purchase@iitpkd.ac.in
---	---	--

CLARIFICATIONS AFTER PRE-BID MEETING

DATE : 03.01.2018

DUE DATE OF THE TENDER: 17.01.2018 AT 2.30 PM

Clarifications for Tender: IITPKD/CIF/JB/089/2017, Date: 22.12.2017

1. Technical Specification, Section 1: [Vacuum Coating Unit].
“Pneumatically operated High Vacuum Gate Valve to isolate Turbo pumps” should be read as
“*Pneumatically operated High Vacuum three position Gate Valve or appropriate throttle valve to isolate Turbo pumps*”
2. Technical Specification, Section 4:
Point iii - the power supply for e-beam should be read as *three-phase power supply*.
Point iii - the e-beam gun should have the sweep controller ability for uniform evaporation of the target material.
Point iv – Crucible indexer should be of an internationally reputed make (e.g. Telemark or substantially equivalent and internationally reputed make)
3. Technical Specification, Section 6: (Quartz control thickness monitor and controller)
Quartz control thickness controller and monitor should of an internationally reputed make (for eg. INFICON or substantially equivalent and internationally reputed make)
4. Technical Specification, Section 12: (Optional Items)
(a) Additional thermocouple should be quoted with accessories for obtaining the temperature of the chamber during deposition.
(b) Alumina coated graphite crucibles

Rest of the terms and conditions of the tender will remain unchanged

REGISTRAR, IIT PALAKKAD